

New England Biolabs Certificate of Analysis


Product Name: WarmStart[®] LAMP Kit (DNA & RNA)
Catalog Number: E1700L
Packaging Lot Number: 10204108
Expiration Date: 07/2025
Storage Temperature: -20°C
Specification Version: PS-E1700S/L v3.0

WarmStart [®] LAMP Kit (DNA & RNA) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M1700SVIAL	WarmStart [®] LAMP 2X Master Mix	10200587	Pass
B1700SVIAL	LAMP Fluorescent Dye	10194694	Pass

Assay Name/Specification	Lot # 10204108
Functional Testing (LAMP, Master Mix) A 25 µl reaction with 1X WarmStart [®] LAMP Master Mix in the presence of 1X LAMP Primers containing 10 ng genomic DNA and 1X LAMP fluorescent dye results in a threshold time of ≤ 20 minutes as determined by fluorescent detection.	Pass
Functional Testing (RT-LAMP, Master Mix) A 25 µl reaction with 1X WarmStart [®] LAMP Master Mix in the presence of 1X LAMP Primers containing 10 ng of genomic RNA and 1X LAMP fluorescent dye results in a threshold time of ≤ 20 minutes as determined by fluorescent detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.


Lea Antonopoulos
Production Scientist
31 Aug 2023


Michael Tonello
Packaging Quality Control Inspector
31 Aug 2023