

New England Biolabs Certificate of Analysis

Product Name: *Tma Endonuclease III*
Catalog Number: *M0291S*
Concentration: *10,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to cleave 1 pmol of a 60-mer oligonucleotide duplex containing a single AP site in a total reaction volume of 10 µl in 1 hour at 65°C.*
Packaging Lot Number: *10154970*
Expiration Date: *06/2024*
Storage Temperature: *-20°C*
Storage Conditions: *10 mM Tris-HCl , 100 mM KCl , 1 mM DTT , 0.1 mM EDTA , 0.1 % Triton®X-100 , 50 % Glycerol, (pH 7.4 @ 25°C)*
Specification Version: *PS-M0291S/L v1.0*

Tma Endonuclease III Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M0291SVIAL	Tma Endonuclease III	10154971	Pass
B9004SVIAL	ThermoPol® Reaction Buffer Pack	10143500	Pass

Assay Name/Specification	Lot # 10154970
Endonuclease Activity (Nicking) A 50 µl reaction in ThermoPol® Reaction Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 100 units of Tma Endonuclease III incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in ThermoPol® Reaction Buffer containing 1 µg of Lambda-HindIII DNA and a minimum of 100 units of Tma Endonuclease III incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 1 containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Tma Endonuclease III incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Lauren Higgins

Lauren Higgins
Production Scientist
15 Jun 2022

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
15 Jun 2022