

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Histone H2A/H2B Dimer Human, Recombinant

Catalog Number: M2508S
Concentration: 20 µM
Unit Definition: N/A

Packaging Lot Number: 10144355
Expiration Date: 06/2023
Storage Temperature: -20°C

Storage Conditions: 2 M NaCl, 20 mM Tris-HCl, 1 mM DTT, 1 mM EDTA, (pH 8.0 @ 25°C)

Specification Version: PS-M2508S v1.0

Histone H2A/H2B Dimer Human, Recombinant Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M2508SVIAL	Histone H2A/H2B Dimer Human, Recombinant	10144356	Pass	

Assay Name/Specification	Lot # 10144355
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in NEBuffer 2 containing 1 μg of a mixture of single and double-stranded [ ³H] E. coli DNA and a minimum of 10 μg of Histone H2A/H2B Dimer Human, Recombinant incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Endonuclease Activity (Nicking) A 50 μl reaction in NEBuffer 2 containing 1 μg of supercoiled PhiX174 RF I DNA and a minimum of 10 μg of Histone H2A/H2B Dimer Human, Recombinant incubated for 4 hours at 37°C results in <10% conversion to RFII as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) Histone H2A/H2B Dimer Human, Recombinant is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
Protease Activity (Histones) A 12 μl reaction containing 7 μl of a standard mixture of proteins and a minimum of 10 μg of Histone H2A/H2B Dimer Human, Recombinant incubated for 4 hours at 37°C, results in no detectable degradation of the protein mixture as determined by SDS-PAGE with Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.



M2508S / Lot: 10144355 Page 1 of 2 One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Fana Mersha Production Scientist

14 Jun 2022

Michael Tonello

Packaging Quality Control Inspector

14 Jun 2022



M2508S / Lot: 10144355

Page 2 of 2