

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® FFPE DNA Repair Mix

Catalog Number: M6630L
Lot Number: 10036818
Expiration Date: 08/2020
Storage Temperature: -20°C

Specification Version: PS-M6630S/L v1.0

NEBNext® FFPE DNA Repair Mix Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M6630LVIAL	NEBNext® FFPE DNA Repair Mix	10036819	Pass	
E6622AAVIAL	NEBNext® FFPE DNA Repair Buffer	10036820	Pass	

Assay Name/Specification	Lot # 10036818
Functional Testing (Oligonucleotide Cleavage - Uracil) A 10 μl reaction in ThermoPol® Reaction Buffer containing 2.5 pmol of annealed oligo containing uracil as the non-standard base and 1 μl of the NEBNext® FFPE DNA Repair Mix incubated for 10 minutes at 37°C resulted in >70% cleavage as determined by polyacrylamide gel electrophoresis.	Pass
Functional Testing (Oligonucleotide Cleavage - Thymine Glycol) A 10 μl reaction in ThermoPol® Reaction Buffer containing 2.5 pmol of annealed oligo containing thymine glycol as the non-standard base and 1 μl of the NEBNext® FFPE DNA Repair Mix incubated for 20 minutes at 37°C resulted in >70% cleavage as determined by polyacrylamide gel electrophoresis.	Pass
PCR Amplification (1 kb) A 48 μl reaction in ThermoPol® Reaction Buffer containing 1.5 ng of UV damaged Lambda DNA, 100 μM dNTPs, 500 μM NAD+ and 1 μl of the NEBNext® FFPE DNA Repair Mix was incubated for 15 minutes at 37°C. Addition of 100 μM dNTPs, 0.4 μM L1 primer mix and 2.5 units of Taq DNA Polymerase followed by 25 cycles of PCR resulted in the expected 1 kb specific product.	Pass
Functional Testing (Oligonucleotide Cleavage - 8-oxo-guanine) A 10 µl reaction in ThermoPol® Reaction Buffer containing 2.5 pmol of annealed oligo containing 8-oxo-guanine as the non-standard base and 1 µl of the NEBNext® FFPE DNA Repair Mix incubated for 1 hour at 37°C resulted in >70% cleavage as determined by polyacrylamide gel electrophoresis.	Pass



M6630L / Lot: 10036818

Page 1 of 2

Lot # 10036818
Pass

This product has been tested and shown to be in compliance with all specifications.

Christine Sumner

Production Scientist 10 Jun 2019 Michael Tonello

Packaging Quality Control Inspector

10 Jun 2019

