

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Nt.BbvCl
Catalog Number: R0632L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to convert 1 µg

of supercoiled pUB DNA to open circular form in 1 hour at 37°C in a

total reaction volume of 50 μl.

Packaging Lot Number: 10156950
Expiration Date: 07/2024
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0632S/L v1.0

Nt.BbvCl Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0632LVIAL	Nt.BbvCl	10156949	Pass	
B6004SVIAL	rCutSmart™ Buffer	10153335	Pass	

Assay Name/Specification	Lot # 10156950
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 30 units of Nt.BbvCl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 30 Units of Nt.BbvCl incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Endonuclease Activity (Nicking) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of supercoiled pUC19 DNA and a minimum of 30 units of Nt.BbvCl incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



R0632L / Lot: 10156950 Page 1 of 2



Pengha Zhang Production Scientist 12 Jul 2022 Michael Tonello

Packaging Quality Control Inspector 12 Jul 2022