

New England Biolabs Certificate of Analysis

Product Name: Q5® High-Fidelity PCR Kit
Catalog Number: E0555L
Packaging Lot Number: 10068693
Expiration Date: 10/2021
Storage Temperature: -20°C
Specification Version: PS-E0555S/L v3.0

| Q5® High-Fidelity PCR Kit Component List | | | |
|--|---|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| N0550AAVIAL | Quick-Load® Purple 1 kb Plus DNA Ladder | 10062317 | Pass |
| M0492SVIAL | Q5® High-Fidelity 2X Master Mix | 10051982 | Pass |
| B7024SVIAL | Gel Loading Dye, Purple (6X) | 10059229 | Pass |
| B1502AVIAL | Nuclease-free Water | 10059286 | Pass |

| Assay Name/Specification | Lot # 10068693 |
|---|----------------|
| <p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in Q5® High-Fidelity PCR Kit and meet the designated specifications.</p> | Pass |
| <p>PCR Amplification (20 kb Lambda DNA, Master Mix) A 50 µl reaction in 1X Q5® High-Fidelity Master Mix and 1.0 µM primers containing 10 ng Lambda DNA for 22 cycles of PCR amplification results in the expected 20 kb product.</p> | Pass |
| <p>PCR Amplification (7 kb Human Genomic DNA, Master Mix) A 50 µl reaction in 1X Q5® High-Fidelity Master Mix and 0.5 µM primers containing 20 ng Human Genomic DNA for 30 cycles of PCR amplification results in the expected 7 kb product.</p> | Pass |

This product has been tested and shown to be in compliance with all specifications.

Christie Vazquez

Christie Vazquez
Production Scientist
07 Oct 2019

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
24 Mar 2020