

## New England Biolabs Certificate of Analysis


Product Name: NEBridge<sup>®</sup> Golden Gate Assembly Kit (Bsal-HF-v2)  
 Catalog Number: E1601L  
 Packaging Lot Number: 10202396  
 Expiration Date: 06/2025  
 Storage Temperature: -20°C  
 Specification Version: PS-E1601S/L v3.0

NEBridge <sup>®</sup> Golden Gate Assembly Kit (Bsal-HF-v2) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
N0309AAVIAL	pGGAselect Destination Plasmid	10195891	Pass
M2616AAVIAL	NEBridge <sup>®</sup> Golden Gate Enzyme Mix (Bsal-HF <sup>®</sup> v2)	10195893	Pass
B0202SVIAL	T4 DNA Ligase Reaction Buffer	10191186	Pass

Assay Name/Specification	Lot # 10202396
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBridge<sup>®</sup> Golden Gate Assembly Kit (Bsal-HFv2) and meet the designated specifications.</p> <p><b>Functional Testing (Golden Gate Assembly)</b>            A 20 µl reaction containing 1X T4 DNA Ligase Reaction Buffer, 75 ng pGGAselect (Golden Gate destination plasmid, CamR), 75 ng each of 5 plasmids carrying fragments of a gene encoding lacZ, and 1 µl NEBridge<sup>®</sup> Golden Gate Assembly Kit (Bsal-HFv2) is incubated for 30 cycles of 37°C for 1 minute, 16°C for 1 minute, and then at 60°C for 5 minutes to linearize any remaining plasmid. Successfully assembled fragments result in lacZ gene in the pGGAselect vector and yield blue colonies on an IPTG/Xgal/Chloramphenicol plate. Transformation of T7 Express Competent E. coli (High Efficiency, NEB #C2566) with 2 µl of the assembly reaction yields &gt;250 colonies.</p>	<p>Pass</p> <p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Lindsey Spiegelman  
Production Scientist  
03 Jul 2023



Josh Hersey  
Packaging Quality Control Inspector  
31 Jul 2023