

New England Biolabs Certificate of Analysis

Product Name: NEBNext[®] Ultra[™] DNA Library Prep Kit for Illumina[®]
 Catalog Number: E7370L
 Packaging Lot Number: 10064024
 Expiration Date: 06/2021
 Storage Temperature: -20°C
 Specification Version: PS-E7370S/L v1.0

| NEBNext [®] Ultra [™] DNA Library Prep Kit for Illumina [®] Component List | | | |
|---|--|------------|----------------------|
| NEB Part Number | Component Description | Lot Number | Individual QC Result |
| E7374AAVIAL | NEBNext [®] Ligation Enhancer | 10059732 | Pass |
| E7373AAVIAL | Blunt/TA Ligase Master Mix | 10059731 | Pass |
| E7372AAVIAL | NEBNext [®] End Repair Reaction Buffer | 10059730 | Pass |
| E7371AAVIAL | NEBNext [®] End Prep Enzyme Mix | 10059729 | Pass |
| E6625AAVIAL | NEBNext [®] Q5 [®] Hot Start HiFi PCR Master Mix | 10059733 | Pass |

| Assay Name/Specification | Lot # 10064024 |
|---|-------------------------|
| <p>* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext[®] Ultra[™] DNA Library Prep Kit for Illumina[®] and meet the designated specifications.</p> <p>Functional Testing (Library Construction, DNA) Each set of reagents is functionally validated and compared to a previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. Libraries made from previous and current lots for both input DNA amounts are sequenced together on the same Illumina[®] flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p> | <p>Pass</p> <p>Pass</p> |

This product has been tested and shown to be in compliance with all specifications.

Christine Sumner

Christine Sumner
Production Scientist
30 Dec 2019

Michael Tonello

Michael Tonello
Packaging Quality Control Inspector
30 Dec 2019