

be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	NEBNext® rRNA Depletion Kit (Bacteria)
Catalog Number:	E7850L
Lot Number:	10056735
Expiration Date:	05/2021
Storage Temperature:	-20°C
Specification Version:	PS-E7850S/L/X v1.0

NEBNext® rRNA Depletion Kit (Bacteria) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7851-3VIAL	NEBNext® Bacterial rRNA Depletion Solution	10054868	Pass
E7753-3VIAL	NEBNext® DNase I	10046657	Pass
E7752-3VIAL	NEBNext® Thermostable RNase H	10046658	Pass
E6317-3VIAL	Nuclease-free Water	10046662	Pass
E6315-3VIAL	DNase I Reaction Buffer	10046661	Pass
E6314-3VIAL	NEBNext® Probe Hybridization Buffer	10046663	Pass
E6312-3VIAL	RNase H Reaction Buffer	10046660	Pass

Assay Name/Specification	Lot # 10056735
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in NEBNext®	
rRNA Depletion Kit (Bacteria) and meet the designated specifications.	
Functional Testing (rRNA Depletion)	Pass
NEBNext® rRNA Depletion Kit (Bacteria) is functionally validated using E. coli RNA	
and RNA from a mock bacterial community. After treatment with the kit, a	
high-throughput sequencing library is made and sequenced. Reads from this library	
are classified as being derived from ribosomal RNA based on sequence homology. This	
method produces libraries with <10% rRNA.	

This product has been tested and shown to be in compliance with all specifications.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

notin /

Christine Sumner Production Scientist 27 Sep 2019

Michae ton l

Michael Tonello Packaging Quality Control Inspector 27 Sep 2019

