

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Adenosine-5 Triphosphate (ATP)

Catalog Number: P0756L
Concentration: 10 mM
Unit Definition: N/A

Lot Number: 10039434
Expiration Date: 01/2021
Storage Temperature: -20°C

Storage Conditions: Milli-Q® Water as a sodium salt, (pH 7.0 @ 25°C)

Specification Version: PS-P0756S/L v1.0

Adenosine-5 Triphosphate (ATP) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
P0756SVIAL	Adenosine 5'-Triphosphate (ATP)	10034352	Pass

Assay Name/Specification	Lot # 10039434
Endonuclease Activity (Nicking) A 50 μl reaction in CutSmart® Buffer containing 1 μg of supercoiled PhiX174 DNA and a minimum of 1 mM of ATP incubated for 4 hours at 30°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Phosphatase Activity (pNPP) A 200 µl reaction in 1M Diethanolamine, pH 9.8, 0.5 mM MgCl2 containing 2.5 mM p-Nitrophenyl Phosphate (pNPP) and a minimum of 1 mM of ATP incubated for 4 hours at 37°C yields <0.0001 unit of alkaline phosphatase activity as determined by spectrophotometric analysis.	Pass
Protease Activity (SDS-PAGE) A 20 µl reaction in 1X CutSmart® Buffer containing 24 µg of a standard mixture of proteins and a minimum of 1 mM of ATP incubated for 16 hours at 37°C, results in no detectable degradation of the protein mixture as determined by SDS-PAGE with Coomassie Blue detection.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 mM of ATP is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass



P0756L / Lot: 10039434

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Alicia Bielik Production Scientist

determined by agarose gel electrophoresis.

25 Jan 2019

Michael Tonello

Packaging Quality Control Inspector

18 Apr 2019



P0756L / Lot: 10039434

Page 2 of 2