

New England Biolabs Certificate of Analysis

Product Name: Diluent A
Catalog Number: B8001S
Concentration: 1 X Concentrate
Lot Number: 10031030
Expiration Date: 12/2021
Storage Temperature: -20°C
Specification Version: PS-B8001S v1.0
Composition (1X): 10 mM Tris-HCl , 50 mM KCl , 1 mM DTT , 0.1 mM EDTA , 200 µg/ml BSA , 50 % Glycerol, (pH 7.4 @ 25°C)

Diluent A Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
B8001SVIAL	Diluent A	10030462	Pass

Assay Name/Specification	Lot # 10031030
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart® Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 10 µl of Diluent A incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart® Buffer containing 1 µg of PhiX174-HaeIII DNA and a minimum of 10 µl of Diluent A incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
pH (buffers/solutions) The pH of 1X Diluent A is between pH 7.3 and 7.5 at 25°C.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 1 µl of Diluent A is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is ≤ 1 E. coli genome.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	Pass

Assay Name/Specification	Lot # 10031030
and a minimum of 1 µl of Diluent A is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
07 Dec 2018



Michael Tonello
Packaging Quality Control Inspector
07 Dec 2018