

## New England Biolabs Certificate of Analysis

Product Name: Luna<sup>®</sup> SARS-CoV-2 RT-qPCR Multiplex Assay Kit  
 Catalog Number: E3019L  
 Packaging Lot Number: 10111841  
 Expiration Date: 03/2022  
 Storage Temperature: -20°C  
 Specification Version: PS-E3019S/L v1.0

Luna <sup>®</sup> SARS-CoV-2 RT-qPCR Multiplex Assay Kit Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S0631AAVIAL	SARS-CoV-2 Primer/Probe Mix (N1/N2/RP)	10083038	Pass
N2117SVIAL	SARS-CoV-2 Positive Control (N gene)	10100711	Pass
M3019SVIAL	Luna <sup>®</sup> Probe One-Step RT-qPCR 4X Mix with UDG	10102378	Pass
B1502AVIAL	Nuclease-free Water	10104354	Pass

Assay Name/Specification	Lot # 10111841
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in Luna<sup>®</sup> SARS-CoV-2 RT-qPCR Multiplex Assay Kit and meet the designated specifications.</p>	Pass
<p><b>Functional Testing (RT-qPCR, SARS-CoV-2)</b>            Luna<sup>®</sup> SARS-CoV-2 RT-qPCR Multiplex Assay Kit is functionally tested and compared to a control lot using 10 ng of human total RNA spiked with synthetic SARS-CoV-2 RNA in an RT-qPCR assay, resulting in a standard curve with a calculated qPCR efficiency of 90-110%, and a dynamic range of 5 orders of magnitude.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christie Vazquez*

Christie Vazquez  
Production Scientist  
01 Jun 2021

*Michael Tonello*

Michael Tonello  
Packaging Quality Control Inspector  
01 Jun 2021