

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> End Repair Module  
 Catalog Number: E6050L  
 Packaging Lot Number: 10146344  
 Expiration Date: 10/2023  
 Storage Temperature: -20°C  
 Specification Version: PS-E6050S/L v1.0

NEBNext <sup>®</sup> End Repair Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E6052AAVIAL	NEBNext <sup>®</sup> End Repair Reaction Buffer	10122395	Pass
E6051AAVIAL	NEBNext <sup>®</sup> End Repair Enzyme Mix	10122394	Pass

Assay Name/Specification	Lot # 10146344
<b>Functional Testing (Library Construction, DNA Modules)</b> Each set of reagents is functionally validated and compared to a previous lot through construction of DNA libraries. Libraries made from previous and current lots for the recommended input DNA amounts are sequenced together on the same Illumina <sup>®</sup> flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	Pass
<b>* Individual Product Component Note</b> Standard Quality Control Tests are performed for each component included in NEBNext <sup>®</sup> End Repair Module and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
23 Mar 2022

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
23 Mar 2022