

## New England Biolabs Certificate of Analysis

**Product Name:** NEBNext<sup>®</sup> Multiplex Oligos for Illumina<sup>®</sup> (96 Unique Dual Index Primer Pairs Set 2)  
**Catalog Number:** E6442L  
**Packaging Lot Number:** 10125866  
**Expiration Date:** 11/2023  
**Storage Temperature:** -20°C  
**Specification Version:** PS-E6442S/L v1.0

NEBNext <sup>®</sup> Multiplex Oligos for Illumina <sup>®</sup> (96 Unique Dual Index Primer Pairs Set 2) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E6612AVIAL	NEBNext <sup>®</sup> Adaptor for Illumina <sup>®</sup>	10125869	Pass
E6610AAVIAL	USER Enzyme	10125868	Pass
E6443AVIAL	NEBNext <sup>®</sup> Unique Dual Index Primer Set 2 Plate	10125867	Pass

Assay Name/Specification	Lot # 10125866
<p><b>Functional Testing (Library Construction, Oligo)</b>            Each of the components are functionally validated through construction of libraries made from commercially available genomic DNA. Libraries are made from each unique barcoded oligonucleotide and are sequenced together on the same Illumina<sup>®</sup> flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.</p> <p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> Multiplex Oligos for Illumina<sup>®</sup> (96 Unique Dual Index Primer Pairs Set 2) and meet the designated specifications.</p>	<p>Pass</p> <p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
21 Dec 2021

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
12 Jan 2022