


New England Biolabs Certificate of Analysis

Product Name: MspI
Catalog Number: R0106L
Concentration: 20,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10061359
Expiration Date: 12/2021
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0106S/L v1.0

MspI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0106LVIAL	MspI	10061360	Pass
B7204SVIAL	CutSmart® Buffer	10055736	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10053980	Pass

Assay Name/Specification	Lot # 10061359
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of MspI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with MspI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with MspI.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of MspI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.



Mala Samaranayake
Production Scientist
06 Dec 2019



Michael Tonello
Packaging Quality Control Inspector
06 Dec 2019