

## New England Biolabs Certificate of Analysis

**Product Name:** *NheI*  
**Catalog Number:** *R0131S*  
**Concentration:** *10,000 U/ml*  
**Unit Definition:** *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (HindIII digest) in 1 hour at 37°C in a total reaction volume of 50 µl.*  
**Lot Number:** *10015942*  
**Expiration Date:** *07/2020*  
**Storage Temperature:** *-20°C*  
**Storage Conditions:** *250 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200 µg/ml BSA*  
**Specification Version:** *PS-R0131S/L v1.0*

NheI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0131SVIAL	NheI	10015941	Pass
B7202SVIAL	NEBuffer™ 2.1	0261805	Pass
B7024SVIAL	Gel Loading Dye, Purple (6X)	10011266	Pass


Assay Name/Specification	Lot # 10015942
<b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in NEBuffer 2.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 10 Units of NheI incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in NEBuffer 2.1 containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 250 units of NheI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 20-fold over-digestion of Lambda HindIII DNA with NheI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with NheI.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in NEBuffer 2.1 containing 1 µg of Lambda HindIII DNA and a minimum	Pass

Assay Name/Specification	Lot # 10015942
of 50 Units of NheI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso  
Production Scientist  
06 Jun 2018



Mary Conlon  
Packaging Quality Control Inspector  
20 Jul 2018