

New England Biolabs Certificate of Analysis

Product Name: BspHI
Catalog Number: R0517S
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10144580
Expiration Date: 03/2024
Storage Temperature: -20°C
Storage Conditions: 100 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0517S/L v1.0

BspHI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0517SVIAL	BspHI	10144579	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10138405	Pass
B6004SVIAL	rCutSmart™ Buffer	10143286	Pass

Assay Name/Specification	Lot # 10144580
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of BspHI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with BspHI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with BspHI.</p>	Pass
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 50 Units of BspHI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass
<p>Protein Purity Assay (SDS-PAGE) BspHI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Penghua Zhang
Production Scientist
04 Apr 2022



Michael Tonello
Packaging Quality Control Inspector
04 Apr 2022