

New England Biolabs Certificate of Analysis


Product Name: Nb.BssSI
Catalog #: R0681S
Concentration: 20,000 units/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pUC19 DNA in NEBuffer 3.1 incubated for 1 hour at 37°C in a total reaction volume of 50 µl.
Lot #: 0051605
Assay Date: 05/2016
Expiration Date: 05/2018
Storage Temp: -20°C
Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl, 1 mM DTT, 0.1 mM EDTA, 50 % Glycerol, 500 µg/ml BSA, (pH 7.4 @ 25°C)
Specification Version: PS-R0681S v1.0
Effective Date: 31 Mar 2016

Assay Name/Specification (minimum release criteria)	Lot #0051605
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 200 units of Nb.BssSI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Measured Activity (Restriction Endonuclease) - The measured activity of Nb.BssSI is complete at 20,000 units/ml and incomplete at 40,000 units/ml.	Pass
Non-Specific DNase Activity (16 hour) - A 50 µl reaction in NEBuffer 3.1 containing 1 µg of pUC19 DNA and a minimum of 20 units of Nb.BssSI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.	Pass
Protein Purity Assay (SDS-PAGE) - Nb.BssSI is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Authorized by
Derek Robinson
31 Mar 2016



Inspected by
Penghua Zhang
06 May 2016

