


New England Biolabs Certificate of Analysis

Product Name: Hydrophilic Streptavidin Magnetic Beads
Catalog Number: S1421S
Concentration: 4 mg/ml
Lot Number: 10055149
Expiration Date: 06/2022
Storage Temperature: 4°C
Storage Conditions: 0.02 % NaN₃, 0.1 % BSA, 0.05 % Tween®20, 1 X PBS, (pH 7.4 @ 25°C)
Specification Version: PS-S1421S v1.0

Hydrophilic Streptavidin Magnetic Beads Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
S1421SVIAL	Hydrophilic Streptavidin Magnetic Beads	10045963	Pass

Assay Name/Specification	Lot # 10055149
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in Hydrophilic Streptavidin Magnetic Bead Storage Buffer containing 1 µg of PhiX174-HaeIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
RNase Activity (Buffer) A 10 µl reaction in Hydrophilic Streptavidin Magnetic Bead Storage Buffer containing 40 ng of a 300 base single-stranded RNA is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by fluorescent detection.	Pass
Functional Binding Assay (Qualitative) Hydrophilic Streptavidin Magnetic Beads (500 µg) were equilibrated and incubated with 200 µl of Biotin Mouse Anti-Human IgG then washed and incubated with 500 µl Human Serum IgG for 1 hour at 25°C, then washed, eluted and evaluated by Tris-Glycine gel to confirm low non-specific binding of extract proteins and high isolation of target.	Pass
Binding Capacity (Magnetic Beads) Hydrophilic Streptavidin Magnetic Beads (500 µg) were equilibrated and incubated with 100 µl of 5 µM 5'-Biotin-dT25-FAM-3' for 1 hour at 25°C. Binding capacity was determined to be >400 pmol of oligo per mg of beads.	Pass

This product has been tested and shown to be in compliance with all specifications.



Brad Landgraf
Production Scientist
28 Jun 2019



Michael Tonello
Packaging Quality Control Inspector
10 Oct 2019