


New England Biolabs Certificate of Analysis

Product Name: Monarch RNA Purification Columns
 Catalog Number: T2007L
 Packaging Lot Number: 10053941
 Expiration Date: 10/2021
 Storage Temperature: 25°C
 Specification Version: PS-T2007L v1.0

Monarch RNA Purification Columns Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
T2018-1	Monarch® Collection Tubes II (50)	10054029	Pass
T2007-1	Monarch RNA Purification Columns	10054004	Pass

Assay Name/Specification	Lot # 10053941
Functional Testing (DNA Recovery and Purity) Twenty-four Monarch® RNA Purification Columns are tested with 10 µg of Lambda/BstEII Marker resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	Pass

This product has been tested and shown to be in compliance with all specifications.



Ana Egana
Production Scientist
30 Apr 2020



Heidi Church
Packaging Quality Control Inspector
30 Apr 2020