


## New England Biolabs Certificate of Analysis

Product Name: *Monarch gDNA Purification Columns*  
 Catalog Number: *T3017L*  
 Lot Number: *10028547*  
 Expiration Date: *06/2020*  
 Storage Temperature: *25°C*  
 Specification Version: *PS-T3017L v1.0*

Assay Name/Specification	Lot # 10028547
<b>Functional Testing (DNA Recovery and Purity)</b> Twenty-four Monarch® gDNA Purification Columns are tested with 5 µg of HindIII digested Lambda resulting in ≥75% recovery in ≥80% of the samples. OD 260/280 and 260/230 are ≥1.75 in ≥80% of the samples.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.



Ana Egana  
Production Scientist  
25 Apr 2019



Heidi Church  
Packaging Quality Control Inspector  
25 Apr 2019