

New England Biolabs Product Specification

<i>Product Name:</i>	<i>Sau3AI</i>
<i>Catalog #:</i>	<i>R0169S/L</i>
<i>Concentration:</i>	<i>5,000 units/ml</i>
<i>Unit Definition:</i>	<i>One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.</i>
<i>Shelf Life:</i>	<i>12 months</i>
<i>Storage Temp:</i>	<i>-20 °C</i>
<i>Storage Conditions:</i>	<i>50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA</i>
<i>Specification Version:</i>	<i>PS-R0169S/L v1.0</i>
<i>Effective Date:</i>	<i>01 May 2013</i>

Assay Name/Specification (minimum release criteria)

Endonuclease Activity (Nicking) - A 50 µl reaction in NEBuffer 1.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 5 Units of Sau3AI incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.

Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in NEBuffer 1.1 containing 1 µg of a mixture of single and double-stranded [³H] *E. coli* DNA and a minimum of 15 units of Sau3AI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.

Ligation and Recutting (Terminal Integrity) - After a 20-fold over-digestion of Lambda DNA with Sau3AI, ~75% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Sau3AI.

Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in NEBuffer 1.1 containing 1 µg of Lambda DNA and a minimum of 25 Units of Sau3AI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.

Protein Purity Assay (SDS-PAGE) - Sau3AI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Date 01 May 2013

Derek Robinson
Director of Quality Control

