

be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	LunaScript® Multiplex One-Step RT-PCR Kit
Catalog Number:	E1555L
Packaging Lot Number:	10157369
Expiration Date:	08/2023
Storage Temperature:	-20°C
Specification Version:	PS-E1555S/L v1.0

LunaScript® Multiplex One-Step RT-PCR Kit Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M1557LVIAL	LunaScript® Multiplex One-Step RT-PCR Reaction Mix	10120305	Pass	
M1556LVIAL	LunaScript® Multiplex One-Step RT-PCR Enzyme Mix	10147794	Pass	
B1502AVIAL	Nuclease-free Water	10151182	Pass	

Assay Name/Specification	Lot # 10157369
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in LunaScript® Multiplex One-Step RT-PCR Kit and meet the designated specifications.	Pass
Functional Testing (RT-PCR, Multiplexing) A 25 µl reaction in 1X LunaScript® Multiplex One-Step RT-PCR Reaction Mix containing 0.1 ng of total RNA, 1X LunaScript® Multiplex One-Step RT-PCR Enzyme Mix, and 0.33 µM primers for 40 cycles of PCR amplification results in the expected three products.	Pass
Functional Testing (RT-PCR, Sensitivity) A 25 µl reaction in 1X LunaScript® Multiplex One-Step RT-PCR Reaction Mix containing 1 pg of total RNA, 1X LunaScript® Multiplex One-Step RT-PCR Enzyme Mix, and 0.5 µM primers for 40 cycles of PCR amplification results in the expected 566 bp product.	Pass

This product has been tested and shown to be in compliance with all specifications.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

vistie Vayquez

Christie Vazquez Production Scientist 18 Jul 2022

Michae 110

Michael Tonello Packaging Quality Control Inspector 18 Jul 2022

