

be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	LunaScript® Multiplex One-Step RT-PCR Ki
Catalog Number:	E1555S
Packaging Lot Number:	10153857
Expiration Date:	10/2023
Storage Temperature:	-20°C
Specification Version:	PS-E1555S/L v1.0

LunaScript® Multiplex One-Step RT-PCR Ki Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
M1557SVIAL	LunaScript® Multiplex One-Step RT-PCR Reaction Mix	10148793	Pass
M1556SVIAL	LunaScript® Multiplex One-Step RT-PCR Enzyme Mix	10148792	Pass
B1502AVIAL	Nuclease-free Water	10145291	Pass

Assay Name/Specification	Lot # 10153857
Functional Testing (RT-PCR, Sensitivity) A 25 µl reaction in 1X LunaScript® Multiplex One-Step RT-PCR Reaction Mix containing 1 pg of total RNA, 1X LunaScript® Multiplex One-Step RT-PCR Enzyme Mix, and 0.5 µM primers for 40 cycles of PCR amplification results in the expected 566 bp product.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in LunaScript® Multiplex One-Step RT-PCR Kit and meet the designated specifications.	Pass
Functional Testing (RT-PCR, Multiplexing) A 25 µl reaction in 1X LunaScript® Multiplex One-Step RT-PCR Reaction Mix containing 0.1 ng of total RNA, 1X LunaScript® Multiplex One-Step RT-PCR Enzyme Mix, and 0.33 µM primers for 40 cycles of PCR amplification results in the expected three products.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

poistie Vayquez

Christie Vazquez Production Scientist 27 May 2022

Michae The b

Michael Tonello Packaging Quality Control Inspector 27 May 2022

