

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: HiScribe™ T7 High Yield RNA Synthesis Kit

Catalog Number: E2040S
Packaging Lot Number: 10109077
Expiration Date: 11/2022
Storage Temperature: -20°C

Specification Version: PS-E2040S v1.0

HiScribe™ T7 High Yield RNA Synthesis Kit Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N0454AVIAL	CTP Solution	10104832	Pass	
N0453AVIAL	UTP Solution	10104831	Pass	
N0452AVIAL	GTP Solution	10104830	Pass	
N0451AVIAL	ATP Solution	10104829	Pass	
N0426AVIAL	Fluc Control Template	10094677	Pass	
M0255AVIAL	T7 RNA Polymerase Mix	10105809	Pass	
B2041AVIAL	10X T7 Reaction Buffer	10098193	Pass	

Assay Name/Specification	Lot # 10109077
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in	
HiScribe™ T7 High Yield RNA Synthesis Kit and meet the designated specifications.	
Functional Testing (RNA Synthesis, Yield)	Pass
A 20 µl reaction in 1X T7 Reaction Buffer in the presence of 40 mM NTPs containing 1	
μg FLuc Control Template with 2 μl of the T7 RNA Polymerase Mix incubated for 2	
hours at 37°C results in a minimum of 150 µg of a 1.8 kb RNA as determined by OD 260	
and agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E2040S / Lot: 10109077

Page 1 of 2

Dongxian Yue
Production Scientist

03 May 2021

Michael Tonello

Packaging Quality Control Inspector

06 May 2021