

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® rRNA Depletion Kit (Human/Mouse/Rat)

Catalog Number: E6310S
Lot Number: 10051231
Expiration Date: 06/2021
Storage Temperature: -20°C

Specification Version: PS-E6310S/L/X/B v1.0

NEBNext® rRNA Depletion Kit (Human/Mouse/Rat) Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
E6318-2VIAL	NEBNext® RNase H	10043007	Pass	
E6317-2VIAL	Nuclease-free Water	10043006	Pass	
E6316-2VIAL	DNase I (RNase-free)	10043005	Pass	
E6315-2VIAL	DNase I Reaction Buffer	10043004	Pass	
E6314-2VIAL	NEBNext® Probe Hybridization Buffer	10043003	Pass	
E6313-2VIAL	NEBNext® rRNA Depletion Solution	10043002	Pass	
E6312-2VIAL	RNase H Reaction Buffer	10043001	Pass	

Assay Name/Specification	Lot # 10051231
Functional Testing (rRNA Depletion) The NEBNext rRNA Depletion Kit is functionally validated using commercially available human RNA (e.g. UHRR). After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.



E6310S / Lot: 10051231

Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Christin Summ

Christine Sumner Production Scientist 26 Jul 2019 Michael Tonello

Packaging Quality Control Inspector

26 Jul 2019

E6310S / Lot: 10051231

Page 2 of 2