

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra II DNA PCR-Free Library Prep Kit for Illumina®

Catalog Number: E7410G
Packaging Lot Number: 10156462
Expiration Date: 06/2023
Storage Temperature: -20°C

Specification Version: PS-E7410S/L/G v1.0

NEBNext® Ultra II DNA PCR-Free Library Prep Kit for Illumina® Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
E7648GVIAL	NEBNext® Ultra™ II Ligation Master Mix	10149018	Pass	
E7647GVIAL	NEBNext® Ultra™ II End Prep Reaction Buffer	10149016	Pass	
E7646GVIAL	NEBNext® Ultra™ II End Prep Enzyme Mix	10149015	Pass	
E7374GVIAL	NEBNext® Ligation Enhancer	10149017	Pass	

Assay Name/Specification	Lot # 10156462
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in NEBNext®	
Ultra™ II DNA PCR-free Library Prep Kit for Illumina® and meet the designated	
specifications.	
Functional Testing (Library Construction, DNA)	Pass
Each set of reagents is functionally validated and compared to a previous lot	
through construction of libraries made from commercially available genomic DNA,	
using the kit's minimum and maximum input requirements. Libraries made from previous	
and current lots for both input DNA amounts are sequenced together on the same	
Illumina® flow cell and compared across various metrics including library yield,	
fraction of reads aligning to the reference, GC bias, and insert size.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7410G / Lot: 10156462

Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Christin Summ

Christine Sumner Production Scientist 22 Jun 2022 Michael Tonello

Packaging Quality Control Inspector

22 Jun 2022



E7410G / Lot: 10156462

Page 2 of 2