

## New England Biolabs Certificate of Analysis

**Product Name:** NEBNext<sup>®</sup> Ultra<sup>™</sup> End Repair/dA-Tailing Module  
**Catalog Number:** E7442L  
**Packaging Lot Number:** 10101443  
**Expiration Date:** 03/2022  
**Storage Temperature:** -20°C  
**Specification Version:** PS-E7442S/L v1.0

NEBNext <sup>®</sup> Ultra <sup>™</sup> End Repair/dA-Tailing Module Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7372AAVIAL	NEBNext <sup>®</sup> End Repair Reaction Buffer	10083125	Pass
E7371AAVIAL	NEBNext <sup>®</sup> End Prep Enzyme Mix	10083124	Pass

Assay Name/Specification	Lot # 10101443
<b>Functional Testing (Library Construction, DNA Modules)</b> Each set of reagents is functionally validated and compared to a previous lot through construction of DNA libraries. Libraries made from previous and current lots for the recommended input DNA amounts are sequenced together on the same Illumina <sup>®</sup> flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	<b>Pass</b>
<b>* Individual Product Component Note</b> Standard Quality Control Tests are performed for each component included in NEBNext <sup>®</sup> Ultra <sup>™</sup> End Repair/dA-Tailing Module and meet the designated specifications.	<b>Pass</b>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
25 Feb 2021

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
25 Feb 2021