

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II DNA Library Prep Kit for Illumina®

Catalog Number: E7645S
Packaging Lot Number: 10082985
Expiration Date: 09/2021
Storage Temperature: -20°C

Specification Version: PS-E7645S/L v1.0

NEBNext® Ultra™ II DNA Library Prep Kit for Illumina® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E7649AVIAL	NEBNext® Ultra™ II Q5® Master Mix	10082990	Pass	
E7648AVIAL	NEBNext® Ultra™ II Ligation Master Mix	10082988	Pass	
E7647AVIAL	NEBNext® Ultra™ II End Prep Reaction Buffer	10082987	Pass	
E7646AVIAL	NEBNext® Ultra™ II End Prep Enzyme Mix	10082986	Pass	
E7374AVIAL	NEBNext® Ligation Enhancer	10082989	Pass	

Assay Name/Specification	Lot # 10082985
Functional Testing (Library Construction, DNA)	Pass
Each set of reagents is functionally validated and compared to a previous lot	
through construction of libraries made from commercially available genomic DNA,	
using the kit's minimum and maximum input requirements. Libraries made from previous	
and current lots for both input DNA amounts are sequenced together on the same	
Illumina® flow cell and compared across various metrics including library yield,	
fraction of reads aligning to the reference, GC bias, and insert size.	
* Individual Product Component Note	Pass
Standard Quality Control Tests are performed for each component included in NEBNext®	
Ultra™ II DNA Library Prep Kit for Illumina® and meet the designated specifications.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7645S / Lot: 10082985

Page 1 of 2



240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Christin Summ

Christine Sumner Production Scientist 22 Sep 2020 Michael Tonello

Packaging Quality Control Inspector

08 Jan 2021

E7645S / Lot: 10082985

Page 2 of 2