

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat)

Catalog Number: E7750S
Packaging Lot Number: 10122737
Expiration Date: 11/2022
Storage Temperature: -20°C

Specification Version: PS-E7750S/L/X v1.0

| NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) Component List | | | | |
|---|---|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| E7753-2VIAL | NEBNext® DNase I | 10089753 | Pass | |
| E7752-2VIAL | NEBNext® Thermostable RNase H | 10089754 | Pass | |
| E7751-2VIAL | NEBNext® Globin & rRNA Depletion Solution | 10089755 | Pass | |
| E6317-2VIAL | Nuclease-free Water | 10089758 | Pass | |
| E6315-2VIAL | DNase I Reaction Buffer | 10089757 | Pass | |
| E6314-2VIAL | NEBNext® Probe Hybridization Buffer | 10089759 | Pass | |
| E6312-2VIAL | RNase H Reaction Buffer | 10089756 | Pass | |

| Assay Name/Specification | Lot # 10122737 |
|---|----------------|
| * Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications. | Pass |
| Functional Testing (Globin and rRNA Depletion) The NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) is functionally validated using commercially available human RNA extracted from blood. After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA and less than 5% globin RNA. | Pass |

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7750S / Lot: 10122737 Page 1 of 2 Christin Summ

Christine Sumner Production Scientist 16 Sep 2021 Michael Tonello

Packaging Quality Control Inspector

16 Sep 2021

E7750S / Lot: 10122737

Page 2 of 2