

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat)

Catalog Number: E7750S
Packaging Lot Number: 10192251
Expiration Date: 04/2025
Storage Temperature: -20°C

Specification Version: PS-E7750S/L/X v1.0

NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E7753-2VIAL	NEBNext® DNase I	10180666	Pass	
E7752-2VIAL	NEBNext® Thermostable RNase H	10180667	Pass	
E7751-2VIAL	NEBNext® Globin & rRNA Depletion Solution	10180668	Pass	
E6317-2VIAL	Nuclease-free Water	10180671	Pass	
E6315-2VIAL	DNase I Reaction Buffer	10180670	Pass	
E6314-2VIAL	NEBNext® Probe Hybridization Buffer	10180672	Pass	
E6312-2VIAL	RNase H Reaction Buffer	10180669	Pass	

Assay Name/Specification	Lot # 10192251
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) and meet the designated specifications.	Pass
Functional Testing (Globin and rRNA Depletion) The NEBNext® Globin & rRNA Depletion Kit (Human/Mouse/Rat) is functionally validated using commercially available human RNA extracted from blood. After treatment with the kit, a high throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with less than 10% rRNA and less than 5% globin RNA.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7750S / Lot: 10192251 Page 1 of 2 Christin Summ

Christine Sumner Production Scientist 15 May 2023 Michael Tonello

Packaging Quality Control Inspector

15 May 2023

E7750S / Lot: 10192251

Page 2 of 2