

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II FS DNA Library Prep Kit for Illumina®

Catalog Number: E7805L
Lot Number: 10045151
Expiration Date: 02/2020
Storage Temperature: -20°C

Specification Version: PS-E7805S/L v1.0

NEBNext® Ultra™ II FS DNA Library Prep Kit for Illumina® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E7808AAVIAL	TE Buffer	10037408	Pass	
E7807AAVIAL	NEBNext® Ultra™ II FS Reaction Buffer	10037407	Pass	
E7806AAVIAL	NEBNext® Ultra™ II FS Enzyme Mix	10037406	Pass	
E7649AAVIAL	NEBNext® Ultra™ II Q5® Master Mix	10037405	Pass	
E7648AAVIAL	NEBNext® Ultra™ II Ligation Master Mix	10037403	Pass	
E7374AAVIAL	NEBNext® Ligation Enhancer	10037402	Pass	

Assay Name/Specification	Lot # 10045151
Functional Testing (Library Construction, FS DNA) Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. A fragmentation time of 20 minutes was used to generate an insert size of approximately 200 bp. The final average library size is between 270 and 450 bp as determined by an Agilent Bioanalyzer. Libraries made from the previous and current lots for both input DNA amounts are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II FS DNA Library Prep Kit for Illumina® and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.



E7805L / Lot: 10045151 Page 1 of 2 Christin Summ

Christine Sumner Production Scientist 17 May 2019 Michael Tonello

Packaging Quality Control Inspector

17 May 2019