

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: NEBNext® Ultra™ II FS DNA Library Prep Kit for Illumina®

Catalog Number: E7805S
Lot Number: 10051578
Expiration Date: 05/2020
Storage Temperature: -20°C

Specification Version: PS-E7805S/L v1.0

NEBNext® Ultra™ II FS DNA Library Prep Kit for Illumina® Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E7808AVIAL	TE Buffer	10046468	Pass	
E7807AVIAL	NEBNext® Ultra™ II FS Reaction Buffer	10046467	Pass	
E7806AVIAL	NEBNext® Ultra™ II FS Enzyme Mix	10046466	Pass	
E7649AVIAL	NEBNext® Ultra™ II Q5® Master Mix	10046465	Pass	
E7648AVIAL	NEBNext® Ultra™ II Ligation Master Mix	10046464	Pass	
E7374AVIAL	NEBNext® Ligation Enhancer	10046463	Pass	

Assay Name/Specification	Lot # 10051578
Functional Testing (Library Construction, FS DNA) Each set of reagents is functionally validated and compared to the previous lot through construction of libraries made from commercially available genomic DNA, using the kit's minimum and maximum input requirements. A fragmentation time of 20 minutes was used to generate an insert size of approximately 200 bp. The final average library size is between 270 and 450 bp as determined by an Agilent Bioanalyzer. Libraries made from the previous and current lots for both input DNA amounts are sequenced together on the same Illumina flow cell and compared across various metrics including library yield, fraction of reads aligning to the reference, GC bias, and insert size.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® Ultra™ II FS DNA Library Prep Kit for Illumina® and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.



E7805S / Lot: 10051578

Christin Summ

Christine Sumner Production Scientist 31 Jul 2019 Michael Tonello

Packaging Quality Control Inspector

31 Jul 2019