

be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	Thermostable OGG
Catalog Number:	M0464S
Concentration:	8,000 U/ml
Unit Definition:	One unit is defined as the amount of enzyme required to cleave 10 pmol of a 60-mer fluorescently labeled oligonucleotide duplex containing a single 8-oxoguanine base paired with a cytosine in a total reaction volume of 10 μ l in 1 hour at 65°C.
Packaging Lot Number:	10179018
Expiration Date:	02/2025
Storage Temperature:	-20°C
Storage Conditions:	10 mM Tris-HCl, 500 mM NaCl, 0.1 mM EDTA, 1 mM DTT, 200 μg/ml rAlbumin (pH 7.4 @ 25°C)
Specification Version:	PS-M0464S v1.0

Thermostable OGG Component List					
NEB Part Number	Component Description	Lot Number	Individual QC Result		
M0464SVIAL	Thermostable OGG	10178881	Pass		
B6004SVIAL	rCutSmart™ Buffer	10175290	Pass		

Assay Name/Specification	Lot # 10179018
DNase Activity (Labeled Oligo, 3' extension) A 50 µl reaction in CutSmart® Buffer containing a 20 nM solution of a fluorescent labeled double-stranded oligonucleotide containing a 3' extension and a minimum of 40 units of Thermostable OGG incubated for 16 hours at 37°C yields <5% degradation as determined by capillary electrophoresis.	Pass
DNase Activity (Labeled Oligo, 5' extension) A 50 µl reaction in CutSmart® Buffer containing a 20 nM solution of a fluorescent labeled double-stranded oligonucleotide containing a 5' extension and a minimum of 40 units of Thermostable OGG incubated for 16 hours at 37°C yields <5% degradation as determined by capillary electrophoresis.	Pass
Double Stranded DNase Activity (Labeled Oligo) A 50 µl reaction in CutSmart® Buffer containing a 20 nM solution of a fluorescent labeled double-stranded oligonucleotide containing a blunt end and a minimum of 40 units of Thermostable OGG incubated for 16 hours at 37°C yields <5% degradation as determined by capillary electrophoresis.	Pass





be INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Assay Name/Specification	Lot # 10179018
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 1 containing 1 µg of Lambda-HindIII DNA and a minimum of 40 units of Thermostable OGG incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Protein Purity Assay (SDS-PAGE) Thermostable OGG is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
qPCR DNA Contamination (E. coli Genomic) A minimum of 8 units of Thermostable OGG is screened for the presence of E. coli genomic DNA using SYBR® Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is \leq 1 E. coli genome.	Pass
RNase Activity (Extended Digestion) A 10 μl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 8 units of Thermostable OGG is incubated at 37°C. After incubation for 16 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass
Single Stranded DNase Activity (FAM-Labeled Oligo) A 50 µl reaction in CutSmart® Buffer containing a 20 nM solution of a fluorescent internal labeled oligonucleotide and a minimum of 40 units of Thermostable OGG incubated for 16 hours at 37°C yields <5% degradation as determined by capillary electrophoresis.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

k,

Mala-Samaranayake Production Scientist **/** 27 Jan 2023

Josh Hersey Packaging Quality Control Inspector

02 Feb 2023

