

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTaq® DNA Polymerase

Catalog Number: M0480L Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 15

nmol of dNTP into acid insoluble material in 30 minutes at 75°C.

Packaging Lot Number: 10100705 Expiration Date: 01/2023 Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl , 100 mM KCl , 1 mM DTT , 0.1 mM EDTA , 0.5 % Tween®

20 , 0.5 % IGEPAL® CA-630 , 50 % Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0480S/L/X v2.0

OneTaq® DNA Polymerase Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0480LVIAL	OneTaq® DNA Polymerase	10096418	Pass	
B9026AVIAL	OneTaq® High GC Enhancer	10096250	Pass	
B9023SVIAL	OneTaq® GC Reaction Buffer	10101203	Pass	
B9022SVIAL	OneTaq® Standard Reaction Buffer	10100213	Pass	

Assay Name/Specification	Lot # 10100705
PCR Amplification (5.0 kb Lambda DNA) A 25 μl reaction in OneTaq® Standard Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 5 ng Lambda DNA with 0.625 units of OneTaq® DNA Polymerase for 25 cycles of PCR amplification results in the expected 5.0 kb product.	Pass
PCR Amplification (Buffer Dependent, >65% GC-rich) A 25 μl reaction in OneTaq® GC Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® DNA Polymerase for 30 cycles of PCR amplification results in the buffer-dependent production of the expected 737 bp product.	Pass
PCR Amplification (Enhancer Dependent, >70% GC-rich) A 25 µl reaction in OneTaq® GC Reaction Buffer and 20% OneTaq® High GC Enhancer in the presence of 200 µM dNTPs and 0.2 µM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® DNA Polymerase for 30 cycles of PCR amplification results in the enhancer-dependent production of the expected 627 bp product.	Pass



M0480L / Lot: 10100705

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

gel electrophoresis using fluorescent detection.

vistie Vazguez

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 19 Mar 2021 Michael Tonello

Packaging Quality Control Inspector

19 Mar 2021

