

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTag® Hot Start DNA Polymerase

Catalog Number: M0481L
Concentration: 5,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 15

nmol of dNTP into acid insoluble material in 30 minutes at 75°C.

Lot Number: 10048747
Expiration Date: 03/2021
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl, 100 mM KCl, 1 mM DTT, 0.1 mM EDTA, 0.5 % Tween®

20 , 0.5 % IGEPAL® CA-630 , 50 % Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0481S/L/X v1.0

OneTaq® Hot Start DNA Polymerase Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0481LVIAL	OneTaq® Hot Start DNA Polymerase	10036341	Pass	
B9026AVIAL	OneTaq® High GC Enhancer	10031487	Pass	
B9023SVIAL	OneTaq® GC Reaction Buffer	10034012	Pass	
B9022SVIAL	OneTaq® Standard Reaction Buffer	10034011	Pass	

Assay Name/Specification	Lot # 10048747
Inhibition of Primer Extension (Hot Start, Radioactivity Incorporation) A 50 µl primer extension assay in ThermoPol® Reaction Buffer in the presence of 200 µM dNTPs including [³H]-dTTP, containing 15 nM primed single-stranded M13mp18 with 2.5 units of OneTaq® Hot Start DNA Polymerase incubated for 16 hours at 25°C yields >95% inhibition when compared to a non-hot start control reaction.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 2 containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA and a minimum of 5 units of OneTaq® Hot Start DNA Polymerase incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
PCR Amplification (5.0 kb Lambda DNA) A 25 μl reaction in OneTaq® Standard Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 5 ng Lambda DNA with 0.625 units of OneTaq® Hot Start DNA Polymerase for 25 cycles of PCR amplification results in the expected 5.0 kb product.	Pass



M0481L / Lot: 10048747

Page 1 of 2

Assay Name/Specification	Lot # 10048747
PCR Amplification (Buffer Dependent, >65% GC-rich) A 25 μl reaction in OneTaq® GC Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® Hot Start DNA Polymerase for 30 cycles of PCR amplification results in the buffer-dependent production of the expected 737 bp product.	Pass
PCR Amplification (Enhancer Dependent, >70% GC-rich) A 25 μl reaction in OneTaq® GC Reaction Buffer and 20% OneTaq® High GC Enhancer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 ng Human Genomic DNA with 0.625 units of OneTaq® Hot Start DNA Polymerase for 30 cycles of PCR amplification results in the enhancer-dependent production of the expected 627 bp product.	Pass
PCR Amplification (Hot Start 2 kb Lambda DNA) A 25 μl reaction in OneTaq® Standard Reaction Buffer in the presence of 200 μM dNTPs and 0.2 μM primers containing 10 pg Lambda DNA and 50 ng Human Genomic DNA with 0.625 units of OneTaq® Hot Start DNA Polymerase for 30 cycles of PCR amplification results in an increase in yield of the 2 kb Lambda product and a decrease in	Pass

This product has been tested and shown to be in compliance with all specifications.

non-specific genomic bands when compared to a non-hot start control reaction.

After incubation for 16 hours, >90% of the substrate RNA remains intact as

determined by gel electrophoresis using fluorescent detection.

A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of OneTaq® Hot Start DNA Polymerase is incubated at 37°C.

Christie Vazquez Production Scientist

RNase Activity (Extended Digestion)

06 Mar 2019

Michael Tonello

Packaging Quality Control Inspector

02 Jul 2019



Pass