

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: OneTag® Quick-Load® 2X Master Mix with Standard Buffer

Catalog Number: M0486L

Concentration: 2 X Concentrate

Lot Number: 10010241
Expiration Date: 06/2020
Storage Temperature: -20°C

Specification Version: PS-M0486S/L v2.0

Composition (1X): 20 mM Tris-HCl (pH 8.9 @ 25°C), 22 mM KCl, 22 mM NH4Cl, 1.8 mM

MgCl2, 0.2 mM dATP, 0.2 mM dCTP, 0.2 mM dGTP, 0.2 mM dTTP, 5 % Glycerol, 0.06 % IGEPAL® CA-630, 0.05 % Tween® 20, 1 X Xylene cyanol, 1 X Tartrazine, 25 units/ml OneTag® DNA Polymerase

OneTaq® Quick-Load® 2X Master Mix with Standard Buffer Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
M0486SVIAL	OneTaq® Quick-Load® 2X Master Mix with Standard Buffer	10009992	Pass	

Assay Name/Specification	Lot # 10010241
Non-Specific DNase Activity (16 hour, Buffer) A 50 µl reaction in 1X OneTaq® Quick-Load® Master Mix with Standard Buffer containing 1 µg of T3 DNA in addition to a reaction containing Lambda-HindIII DNA incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
PCR Amplification (5 kb Lambda, Master Mix) A 25 µl reaction in 1X OneTaq® Quick-Load® Master Mix with Standard Buffer and 0.2 µM primers containing 5 ng Lambda DNA for 25 cycles of PCR amplification results in the expected 5 kb product.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA and a minimum of 1 µl of OneTaq® Quick-Load® 2X Master Mix with Standard Buffer is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by gel electrophoresis using fluorescent detection.	Pass

This product has been tested and shown to be in compliance with all specifications.



M0486L / Lot: 10010241

Page 1 of 2

17-

Lynne Apone Production Scientist 11 Jun 2018 Michael Tonello

Packaging Quality Control Inspector

12 Jun 2018