

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: Phusion® Hot Start Flex DNA Polymerase

Catalog Number: M0535L
Concentration: 2,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme that will incorporate 10

nmol of dNTP into acid insoluble material in 30 minutes at 74°C.

Packaging Lot Number: 10104082
Expiration Date: 10/2022
Storage Temperature: -20°C

Storage Conditions: 20 mM Tris-HCl , 100 mM KCl , 1 mM DTT , 0.1 mM EDTA , 200 µg/ml BSA

, 1X Stabilizers , 50 % Glycerol, (pH 7.4 @ 25°C)

Specification Version: PS-M0535S/L v1.0

Phusion® Hot Start Flex DNA Polymerase Component List				
<b>NEB Part Number</b>	Component Description	Lot Number	Individual QC Result	
M0535LVIAL	Phusion® Hot Start Flex DNA Polymerase	10086614	Pass	
B0519SVIAL	Phusion® GC Buffer Pack	10085574	Pass	
B0518SVIAL	Phusion® HF Buffer Pack	10096255	Pass	
B0515AVIAL	DMSO	10096249	Pass	
B0510AVIAL	MgCl2 Solution (50 mM)	10096253	Pass	

Assay Name/Specification	Lot # 10104082
PCR Amplification (Hot Start, Human Genomic DNA) A 25 μl reaction in Phusion® GC Buffer in the presence of 200 μM dNTPs and 0.5 μM primers containing 50 ng Human Genomic DNA with 0.5 units of Phusion® Hot Start Flex DNA Polymerase for 25 cycles of PCR amplification results in the expected 665 bp product, and a decrease in non-specific genomic bands after pre-incubation at room temperature for 1 hour, when compared to a non-hot start control reaction.	Pass
PCR Amplification (20 kb Lambda DNA) A 50 μl reaction in Phusion® HF Buffer in the presence of 200 μM dNTPs and 1.0 μM primers containing 10 ng Lambda DNA with 1 unit of Phusion® Hot Start Flex DNA Polymerase for 22 cycles of PCR amplification results in the expected 20 kb product.	Pass
PCR Amplification (7.5 kb Human Genomic DNA) A 50 μl reaction in Phusion® HF Buffer in the presence of 200 μM dNTPs and 1.0 μM primers containing 50 ng Human Genomic DNA with 1 unit of Phusion® Hot Start Flex DNA Polymerase for 30 cycles of PCR amplification results in the expected 7.5 kb	Pass



M0535L / Lot: 10104082 Page 1 of 2 This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Christie Vazquez Production Scientist 07 Apr 2021

ristie Vazguez

Josh Hersey

Packaging Quality Control Inspector

07 Apr 2021

