

*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	Vaccinia Capping System
Catalog Number:	M2080S
Concentration:	10,000 U/ml
Unit Definition:	One unit of Vaccinia Capping Enzyme is defined as the amount of enzyme required to incorporate 10 pmol of (α <sup>32</sup> P) GTP into an 80 nt transcript in 1 hour at 37°C.
Packaging Lot Number:	10185170
Expiration Date:	01/2025
Storage Temperature:	-20°C
Storage Conditions:	100 mM NaCl , 20 mM Tris-HCl (pH 8.0), 1 mM DTT , 0.1 mM EDTA , 50 % Glycerol , 0.1 % Triton®X-100
Specification Version:	PS-M2080S v1.0

Vaccinia Capping System Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N2080AVIAL	GTP	10176039	Pass	
M2080SVIAL	Vaccinia Capping System	10177101	Pass	
B9003SVIAL	S-adenosylmethionine (SAM)	10171171	Pass	
B2080AVIAL	10X Capping Buffer	10182783	Pass	

Assay Name/Specification	Lot # 10185170
<b>Endonuclease Activity (Nicking)</b> A 50 µl reaction in Capping Buffer containing 1 µg of supercoiled PhiX174 DNA and a	Pass
minimum of 10 units of Vaccinia Capping System incubated for 4 hours at 37°C results in <10% conversion to the nicked form as determined by agarose gel electrophoresis.	
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in Capping Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 10 units of Vaccinia Capping System incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Protein Purity Assay (SDS-PAGE) Vaccinia Capping System is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection.	Pass
RNase Activity (Extended Digestion) A 10 µl reaction in NEBuffer 4 containing 40 ng of a 300 base single-stranded RNA	Pass





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Assay Name/Specification	Lot # 10185170
and a minimum of 10 units of Vaccinia Capping System is incubated at 37°C. After incubation for 4 hours, >90% of the substrate RNA remains intact as determined by	
gel electrophoresis using fluorescent detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Jessica Cane Production Scientist 25 Jan 2023

Josh Hersey

Packaging Quality Control Inspector 23 Mar 2023

