240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: pBR322 DNA-BstNI Digest

 Catalog #:
 N3031S/L

 Concentration:
 1 mg/ml

 Lot #:
 0621604

 Assay Date:
 04/2016

 Expiration Date:
 4/2018

 Storage Temp:
 -20°C

Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA

Specification Version: PS-N3031S/L v1.0
Effective Date: 19 Jun 2015

Assay Name/Specification (minimum release criteria)	Lot #0621604
A260/A280 Assay - The ratio of UV absorption of pBR322 DNA-BstNI Digest at 260 and 280 nm is between 1.8 and 2.0.	Pass
DNA Concentration (A260) - The concentration of pBR322 DNA-BstNI Digest is between 1000 and 1050 μg/ml as determined by UV absorption at 260 nm.	Pass
Electrophoretic Pattern (Marker) - The banding pattern of pBR322 DNA-BstNI Digest on a 3% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of $0.5~\mu g/ml$ .	Pass
Non-Specific DNase Activity (DNA, 16 hour) - A 50 μl reaction in 1X NEBuffer 2 containing 5 μg of pBR322 DNA-BstNI Digest incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Authorized by Derek Robinson 19 Jun 2015

nqa.
ISO 9001
Registered
Quality





Inspected by Vanessa Mathieu-Sheltry

12 Apr 2016