

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: 50 bp DNA Ladder

Catalog Number: N3236L
Concentration: 1,000 µg/ml

Unit Definition: N/A

Packaging Lot Number: 10177189
Expiration Date: 01/2025
Storage Temperature: -20°C

Storage Conditions: 10 mM Tris-HCl (pH 8.0), 1 mM EDTA

Specification Version: PS-N3236S/L v1.0

50 bp DNA Ladder Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
N3236LVIAL	50 bp DNA Ladder	10176118	Pass	
B7025SVIAL	Gel Loading Dye, Purple (6X), no SDS	10167589	Pass	

Assay Name/Specification	Lot # 10177189
Non-Specific DNase Activity (DNA, 16 hour) A 50 µl reaction in 1X NEBuffer 2 containing 5 µg of 50 bp DNA Ladder incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Electrophoretic Pattern (Marker) The banding pattern of 50 bp DNA Ladder on a 3% agarose gel shows discrete, clearly identifiable bands at each band of the marker, when stained with Ethidium Bromide at a concentration of 0.5 μg/ml.	Pass
DNA Concentration (A260) The concentration of 50 bp DNA Ladder is between 1000 and 1050 μg/ml as determined by UV absorption at 260 nm.	Pass
A260/A280 Assay The ratio of UV absorption of 50 bp DNA Ladder at 260 and 280 nm is between 1.8 and 2.0.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



N3236L / Lot: 10177189



Vanessa Mathieu-Sheltry Production Scientist 04 Jan 2023 Michael Tonello

Packaging Quality Control Inspector

09 Jan 2023