

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name:	cAMP-dependent Protein Kinase (PKA), catalytic subunit
Catalog #:	P6000S/L
Concentration:	2,500,000 units/ml
Unit Definition:	One unit is defined as the amount of PKA catalytic subunit required to catalyze the transfer of 1 pmol of phosphate to Kemptide, LRRASLG (100 μ M) in 1 minute at 30°C in a total reaction volume of 25 μ L.
<i>Lot</i> #:	0181707
Assay Date:	07/2017
Expiration Date:	7/2018
Storage Temp:	-20°C
Storage Conditions:	50 mM NaCl , 20 mM Tris-HCl , 2 mM DTT , 1 mM EDTA , 50 % Glycerol, (pH 7.5 @ 25°C)
Specification Version:	PS-P6000S/L v1.0
Effective Date:	19 Feb 2016

Assay Name/Specification (minimum release criteria)	Lot #0181707
Phosphatase Activity (pNPP) - A 220 μl reaction in NEBuffer for Protein Kinases containing 50 mM <i>p</i> - Nitrophenyl Phosphate (pNPP) and a minimum of 20,000 units cAMP-dependent Protein Kinase (PKA), catalytic subunit incubated for 2 hours at 30°C yields no detectable phosphatase activity as determined by spectrophotometric analysis.	Pass
Protease Activity (SDS-PAGE) - A 20 μ l reaction in 1X NEBuffer for Protein Kinases containing 24 μ g of a standard mixture of proteins and a minimum of 20,000 units of cAMP-dependent Protein Kinase (PKA), catalytic subunit incubated for 2 hours at 30°C, results in no detectable degradation of the protein mixture as determined by SDS-PAGE with Coomassie Blue detection.	Pass

otunen

Authorized by Derek Robinson 19 Feb 2016



gral 20,

Inspected by Brad Landgraf 28 Jun 2017