

*be* INSPIRED *drive* DISCOVERY *stay* GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name:	Blue Prestained Protein Standard, Broad Range (11-250 kDa)
Catalog Number:	P7718S
Packaging Lot Number:	10094818
Expiration Date:	01/2023
Storage Temperature:	-20°C
Storage Conditions:	62.5 mM Tris-Phosphate, 1 mM EDTA, 2% SDS, 10 mM DTT, 1 mM NaN3, 33% Glycerol
Specification Version:	PS-P7718S/L v1.0

Blue Prestained Protein Standard, Broad Range (11-250 kDa) Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
P7718SVIAL	Blue Prestained Protein Standard, Broad Range (11-250 kDa)	10094816	Pass

Assay Name/Specification	Lot # 10094818
<b>Electrophoretic Pattern (Protein Marker)</b> The banding pattern of Blue Prestained Protein Standard, Broad Range (11-250 kDa) on a 10-20% Tris-Glycine gel shows discrete, clearly identifiable bands at each band of the marker, as determined by gel electrophoresis using Coomassie Blue detection.	Pass
<b>Protease Activity (Protein Markers &amp; Ladders)</b> A 20 μl sample of Blue Prestained Protein Standard, Broad Range (11-250 kDa)	Pass
incubated for 16 hours at 37°C, results in no detectable degradation of the protein mixture as determined by SDS-PAGE with Coomassie Blue detection.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.





be INSPIRED drive DISCOVERY stay GENUINE

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

Alicia Bielik Production Scientist 04 Feb 2021

Michae MA.

Michael Tonello Packaging Quality Control Inspector 04 Feb 2021

