

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Hgal
Catalog Number: R0154L
Concentration: 2,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of PhiX174 DNA in 1 hour at 37°C in a total reaction volume of 50

μl.

Lot Number: 10034661
Expiration Date: 01/2021
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0154S/L v1.0

| Hgal Component List | | | | |
|------------------------|-----------------------|------------|----------------------|--|
| NEB Part Number | Component Description | Lot Number | Individual QC Result | |
| R0154LVIAL | Hgal | 10034662 | Pass | |
| B7201SVIAL | NEBuffer™ 1.1 | 0131803 | Pass | |

| Assay Name/Specification | Lot # 10034661 |
|--|----------------|
| Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 1.1 containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 2 units of Hgal incubated for 4 hours at 37°C releases <0.2% of the total radioactivity. | Pass |
| Ligation and Recutting (Terminal Integrity) After a 2-fold over-digestion of PhiX174 DNA with Hgal, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Hgal. | Pass |
| Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 1.1 containing 1 µg of PhiX174 DNA and a minimum of 2 Units of Hgal incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass |
| Protein Purity Assay (SDS-PAGE) Hgal is ≥ 95% pure as determined by SDS-PAGE analysis using Coomassie Blue detection. | Pass |



R0154L / Lot: 10034661

Page 1 of 2

NEW ENGLAND BioLabs Inc.

This product has been tested and shown to be in compliance with all specifications.

Penghaa Zhang Production Scientist

23 Jan 2019

Michael Tonello

Packaging Quality Control Inspector

09 Apr 2019

R0154L / Lot: 10034661

Page 2 of 2