

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: DdeI

Catalog #: R0175S/L
Concentration: 10,000 units/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction

volume of 50 μ l.

 Lot #:
 1501501

 Assay Date:
 01/2015

 Expiration Date:
 01/2017

 Storage Temp:
 -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 500 µg/ml BSA

Specification Version: PS-R0175S/L v1.0
Effective Date: 28 Mar 2014

Assay Name/Specification (minimum release criteria)	Lot #1501501
Exonuclease Activity (Radioactivity Release) - A 50 μl reaction in CutSmart TM Buffer containing 1 μg of a mixture of single and double-stranded [³ H] E, coli DNA and a minimum of 30 units of DdeI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 10-fold over-digestion of Lambda DNA with DdeI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with DdeI.	Pass
Non-Specific DNase Activity (16 Hour) - A 50 μ l reaction in CutSmart TM Buffer containing 1 μ g of Lambda DNA and a minimum of 10 Units of DdeI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Authorized by Derek Robinson 28 Mar 2014

nga.
ISO 9001
Registered
Quality





Robert Maunus
Inspected by
Robert Maunus
02 Jan 2015

^{*} The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.