

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: Alwl
Catalog Number: R0513L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA (dam-) in 1 hour at 37°C in total reaction volume of

50 μl.

Packaging Lot Number: 10205099
Expiration Date: 08/2025
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0513S/L v1.0

Alwi Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0513LVIAL	AlwI	10201241	Pass	
B6004SVIAL	rCutSmart™ Buffer	10198644	Pass	

Assay Name/Specification	Lot # 10205099
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 10 units of Alwl incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 2-fold over-digestion of Lambda dam- DNA with AlwI, ~50% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, ~75% can be recut with AlwI.	Pass
Non-Specific DNase Activity (16 hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda dam- DNA and a minimum of 10 Units of AlwI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.	Pass



R0513L / Lot: 10205099

Page 1 of 2



Assay Name/Specification	Lot # 10205099
Protein Purity Assay (SDS-PAGE) AlwI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

YunJie Sun \
Production Scientist

Production Scientist 25 Aug 2023

Michael Tonello

Packaging Quality Control Inspector

29 Aug 2023



R0513L / Lot: 10205099 Page 2 of 2