

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: AlwNI
Catalog Number: R0514L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10153992 Expiration Date: 06/2024 Storage Temperature: -20°C

Storage Conditions: 50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0514S/L v1.0

AlwNI Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0514LVIAL	AlwNI	10153991	Pass	
B6004SVIAL	rCutSmart™ Buffer	10150373	Pass	

Assay Name/Specification	Lot # 10153992
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with AlwNI, >95% of the DNA fragments	Pass
can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with AlwNI.	
Non-Specific DNase Activity (16 Hour)	Pass
A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 10 Units of AlwNl incubated for 16 hours at 37°C results in a DNA pattern free of	
detectable nuclease degradation as determined by agarose gel electrophoresis.	
Exonuclease Activity (Radioactivity Release)	Pass
A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of AlwNI incubated for	
4 hours at 37°C releases <0.1% of the total radioactivity.	
Protein Purity Assay (SDS-PAGE)	Pass
AlwNI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	

This product has been tested and shown to be in compliance with all specifications.



R0514L / Lot: 10153992 Page 1 of 2



One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.

Penghua Zhang

Production Scientist 07 Jun 2022

Erin Varney

Packaging Quality Control Inspector

07 Jun 2022

R0514L / Lot: 10153992

Page 2 of 2