

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: BstBl
Catalog Number: R0519L
Concentration: 20,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 65°C in a total reaction volume of 50 μl.

Packaging Lot Number: 1006304
Expiration Date: 11/2021
Storage Temperature: -20°C

Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 200 µg/ml BSA

Specification Version: PS-R0519S/L v1.0

BstBl Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0519LVIAL	BstBI	10060016	Pass	
B7204SVIAL	CutSmart® Buffer	10061304	Pass	

Assay Name/Specification	Lot # 10063041
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Lambda DNA with BstBI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 4 hours at 25°C. Of these ligated fragments, >95% can be recut with BstBI.	Pass
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of supercoiled pUC19 DNA and a minimum of 20 units of BstBl incubated for 4 hours at 65°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 μl reaction in CutSmart™ Buffer containing 1 μg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 100 units of BstBl incubated for 4 hours at 65°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 20 units of BstBl incubated for 16 hours at 65°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass



R0519L / Lot: 10063041

Page 1 of 2



This product has been tested and shown to be in compliance with all specifications.

Anthony Francis
Production Scientist

22 Nov 2019

Michael Tonello

Packaging Quality Control Inspector

22 Jan 2020

R0519L / Lot: 10063041

Page 2 of 2