

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: BspEl
Catalog Number: R0540L
Concentration: 10,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA (dam) in 1 hour at 37°C in a total reaction volume of

50 μl.

Lot Number: 10040229
Expiration Date: 03/2021
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 500 μg/ml BSA

Specification Version: PS-R0540S/L v1.0

BspEl Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0540LVIAL	BspEl	10040230	Pass	
B7203SVIAL	NEBuffer™ 3.1	10033149	Pass	

Assay Name/Specification	Lot # 10040229
Endonuclease Activity (Nicking)	Pass
A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled PhiX174 DNA and a	
minimum of 10 Units of BspEI incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	
donversion to the nicked form as determined by agarose generatiophoresis.	
Exonuclease Activity (Radioactivity Release)	Pass
A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and	
double-stranded [³H] E. coli DNA and a minimum of 100 units of BspEI incubated for	
4 hours at 37°C releases <0.1% of the total radioactivity.	
Ligation and Recutting (Terminal Integrity)	Pass
After a 10-fold over-digestion of Lambda dam- DNA with BspEI, >95% of the DNA	
fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated	
fragments, >95% can be recut with BspEI.	
Non-Specific DNase Activity (16 Hour)	Pass
A 50 µl reaction in NEBuffer 3.1 containing 1 µg of Lambda dam- DNA and a minimum of	
50 units of BspEI incubated for 16 hours at 37°C results in a DNA pattern free of	



R0540L / Lot: 10040229

Page 1 of 2

This product has been tested and shown to be in compliance with all specifications.

Tony Spear-Alfonso Production Scientist

23 Jan 2019

Michael Tonello

Packaging Quality Control Inspector

29 Mar 2019

