

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

Product Name: BbvCl
Catalog Number: R0601L
Concentration: 2,000 U/ml

Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg

of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 μl.

Packaging Lot Number: 10095173
Expiration Date: 11/2022
Storage Temperature: -20°C

Storage Conditions: 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50%

Glycerol, 500 μg/ml BSA

Specification Version: PS-R0601S/L v1.0

BbvCl Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
R0601LVIAL	BbvCl	10089129	Pass	
B7204SVIAL	CutSmart® Buffer	10092681	Pass	

Assay Name/Specification	Lot # 10095173
Exonuclease Activity (Radioactivity Release)	Pass
A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [3H] E. coli DNA and a minimum of 10 units of BbvCl incubated for 4	
hours at 37°C releases <0.1% of the total radioactivity.	
Ligation and Recutting (Terminal Integrity)	Pass
After a 2-fold over-digestion of Lambda DNA with BbvCI, 95% can be recut with BbvCI.	
Non-Specific DNase Activity (16 hour)	Pass
A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of	1 433
2 Units of BbvCl incubated for 16 hours at 37°C results in a DNA pattern free of	
detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE:	
although no nuclease degradation is detected under these conditions, extended	
incubations and/or high concentrations of this enzyme may result in star activity.	
See the product FAQ for recommended reaction conditions for this enzyme.	

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



R0601L / Lot: 10095173 Page 1 of 2



Penghaa Zhang Production Scientist 19 Jan 2021 Michael Tonello

Packaging Quality Control Inspector

19 Jan 2021